



## XTW Extended Wear Collar

Single Patient Use Only

Rx Only

### IMPORTANT INFORMATION

Please read all instructions, warnings, and precautions before use.

### INTENDED USE


This collar is designed to restrict flexion, extension, lateral bending and axial rotation promoting patient recovery.

### PRECAUTIONS

- This product is to be fitted by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners or care providers involved in the patient's care and the patient.
- Consult your physician (or properly licensed practitioner) immediately if you experience sensation changes, unusual reactions, swelling or prolonged pain while using this product.
- Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.

### SIZING

With patient in NEUTRAL ALIGNMENT, measure the vertical distance from the tip of the chin to the



top of the shoulder (A). Select the correct collar by referring to the size guide below. **Note:** cervical region (head, neck and back) of the patient must be in neutral alignment before a measurement is taken.

#### Adult Sizing

Top of Shoulder to



Collar	Chin Dimension	Circumference
X-Short	≤ 1/4"	
Short	1/4" - 5/8"	all are
Regular	5/8" - 1 1/4"	13" - 21"
Tall	1 1/4" - 1 7/8"	
X-Tall	≥ 1 7/8"	

#### Child and Infant Sizing

Collar	Age	Circumference
Infant Short	1 - 18 months	8" - 12"
Infant Regular	9 - 24 months	11" - 13"
Child Tot	1 - 3 years	11 1/2" - 14"
Child Short	2 - 5 years	12" - 14 1/2"
Child Regular	3 - 6 years	13" - 16"

### DIRECTIONS FOR USE

1. Pre-form collar prior to fitting. To pre-form front panel, roll panel sides inward with both hands.
2. To pre-form back panel, roll panel into a cylinder.



## XTW Extended Wear Collar

Single Patient Use Only

Rx Only

### IMPORTANT INFORMATION

Please read all instructions, warnings, and precautions before use.

### INTENDED USE


This collar is designed to restrict flexion, extension, lateral bending and axial rotation promoting patient recovery.

### PRECAUTIONS

- This product is to be fitted by a physician (or properly licensed practitioner) who is familiar with the purpose for which they are responsible. The physician or practitioner is responsible for providing wearing instructions and precautions to other healthcare practitioners or care providers involved in the patient's care and the patient.
- Consult your physician (or properly licensed practitioner) immediately if you experience sensation changes, unusual reactions, swelling or prolonged pain while using this product.
- Follow the instructions of your physician (or properly licensed practitioner) for length and duration of use.

### SIZING

With patient in NEUTRAL ALIGNMENT, measure the vertical distance from the tip of the chin to the



top of the shoulder (A). Select the correct collar by referring to the size guide below. **Note:** cervical region (head, neck and back) of the patient must be in neutral alignment before a measurement is taken.

#### Adult Sizing

Top of Shoulder to


Collar	Chin Dimension	Circumference
X-Short	≤ 1/4"	
Short	1/4" - 5/8"	all are
Regular	5/8" - 1 1/4"	13" - 21"
Tall	1 1/4" - 1 7/8"	
X-Tall	≥ 1 7/8"	

#### Child and Infant Sizing

Collar	Age	Circumference
Infant Short	1 - 18 months	8" - 12"
Infant Regular	9 - 24 months	11" - 13"
Child Tot	1 - 3 years	11 1/2" - 14"
Child Short	2 - 5 years	12" - 14 1/2"
Child Regular	3 - 6 years	13" - 16"


### DIRECTIONS FOR USE

1. Pre-form collar prior to fitting. To pre-form front panel, roll panel sides inward with both hands.
2. To pre-form back panel, roll panel into a cylinder.







3. To apply patient must be in NEUTRAL ALIGNMENT. Slide the front portion of the collar up the patient's neck and properly position the chin support. **Note:** before applying collar, verify that the proper size is selected.



4. Center the back portion onto the patient's neck. With the front portion properly secured, use the bilateral hook and loop straps and fasten the portions together.



5. Minor adjustments can be made to the straps and the front and back flex tabs to achieve desired comfort.



#### CLEANING INSTRUCTIONS

The collar features three removable pads that can be cleaned with mild soap and water. Air dry flat. Replacement pads are available for use during washings.




**Manufacturer**  
DeRoyal  
200 DeBusk Lane, Powell, TN 37849 U.S.A.  
888.938.7828  
865.938.7828

**Authorized Representative**  
DeRoyal Europe, Ltd.  
Virginia Road, Kells, County Meath Ireland  
353.46.9280100


©2007 DeRoyal. All Rights Reserved.  
Part #0-1807  
Revised 5/07




3. To apply patient must be in NEUTRAL ALIGNMENT. Slide the front portion of the collar up the patient's neck and properly position the chin support. **Note:** before applying collar, verify that the proper size is selected.



4. Center the back portion onto the patient's neck. With the front portion properly secured, use the bilateral hook and loop straps and fasten the portions together.



5. Minor adjustments can be made to the straps and the front and back flex tabs to achieve desired comfort.



#### CLEANING INSTRUCTIONS

The collar features three removable pads that can be cleaned with mild soap and water. Air dry flat. Replacement pads are available for use during washings.



**Manufacturer**  
DeRoyal  
200 DeBusk Lane, Powell, TN 37849 U.S.A.  
888.938.7828  
865.938.7828

**Authorized Representative**  
DeRoyal Europe, Ltd.  
Virginia Road, Kells, County Meath Ireland  
353.46.9280100

©2007 DeRoyal. All Rights Reserved.  
Part #0-1807  
Revised 5/07